


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

SEP 08 2009

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130112	✓				
Name		Bruce Smith Jr	ID#	3522	Date	9-1-09	✓	
A	Agency		Wrangell Police Department		Phone #	(907) 874-3304		
Instrument Location		Wrangell Police Department, 431 Zimovia Highway					✓	
Alco S/N		X 172933	✓	Target Value	.080	✓	High Pressure	850
B	Alco Test Values		.085	✓	.086	✓	1 st Alco 2 nd Alco	
Signature								
(OVER)								

BMB
9/15/09

(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz

Scientific Director

State Breath Alcohol Program


Subscribed and sworn before me this 17th day of Sept, 2009.


(Notary Seal Stamp)

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130112 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

SEPTEMBER 01, 2009

OPERATOR'S NAME:

SMITH JR, BRUCE W

OPERATOR'S NUMBER: 3522

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

CAL

O.L. #: 0123456

DEPT/AGENCY: WR61

CASE/REPORT: 09090000

TEST TYPE: U

ALCO TARGET VALUE: .000 ✓

ALCO S/N: X172933 ✓

--- BREATH ANALYSIS ---

.000 ADJUSTED FOR 30.25 in
ALCO TARGET .000 21:20
BLANK TEST .000 21:21
INTERNAL STANDARD VERIFIED 21:21
ALCO TV 30.25 in .005 ✓ 21:22
BLANK TEST .000 21:22
SUBJECT SAMPLE .000 21:23
BLANK TEST .000 21:24
ALCO TV 30.25 in .006 ✓ 21:24
BLANK TEST .000 21:25

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130112

SEPTEMBER 01, 2009 ✓

TIME 21:18

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 48c

BREATH TUBE: 41c

BAROMETER: 30.25 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

"\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHI
HIJKLMNPOQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~